

## How to configure a Fluke data acquisition system for use with transducers

# Application Note

This application note explains how to configure one of the following Fluke data acquisition products when you want to use it with a transducer (for example, a pressure or humidity transducer):

#### Hydra Series Portable Data Acquisition

- NetDAQ<sup>®</sup> Networked Unit
- 2680 Series Data Acquisition System

First you need to determine what will be coming in to the data logger's or data acquisition system's input connector from the transducer. Typical units are 0 to 10 V, 0 to 5 volts, 0 to 20 mA, or 4 to 20 mA. Your transducer specifications will tell you the unit's input and output.



Figure 1. Fluke Calibration current shunt set, part number 2620A-101.



#### Voltage

Simply connect the high and low wires coming from the pressure transducer to the high and low terminals on the desired channel of the data logger's or data acquisition system's universal input connector.

#### Current

You will need a current shunt to convert the current coming in from the transducer to voltage. Fluke Calibration has a readily available set, part number 2620A-101. These shunts, shown below, have a 10  $\Omega$  rating and are designed to work within the input connector for any current source 0 to 100 mA. However, you can use any shunt of the correct value for your application.

Connect the shunt and then connect the wires into the same terminals, just right on top of the shunt as shown in Figure 2.



Figure 2: Setup for current, using a shunt.





### **Software configuration**

Now you'll use the software to configure the channel.

Click the "Channels" button.

featurement Rate	Tempera	ture Units-	OK
Slow (High Res.)	C Celoi	8	Cance
rau	S. Cane	aries.	Help
Scan Parameters		Commu	nication
ligger Type:		(* Dre	d Modilu
Interval Scanning		C Mod	Jem
Scan Interval (secs)	_		Channels
	3	-	
Advanced			Dala File
lydra Front Panel	10	_	0
Monitor Channel No	ne •	ξi -	
		-	

Highlight the channel you want to configure and then click "Function."

hannels:			ÛK.
0 (Ch0 label), OFF 1 (Ch1 label), VDC 3V 2 (Ch2 label), VDC 3V 3 (Ch3 label), VDC 3V	Ĵ.	Function.	Cancel Help
4(04) solen) (CFF 5(D5) solen) (CFF 5(D6) solen) (CFF 5(D6) solen) (CFF 9(D6) solen) (CFF 9(D6) solen) (CFF 9(D6) solen) (CFF 9(D6) solen) (CFF 1(D7) 1 solen) (CFF	a	Alarma Chusmiel Labet [Ch4 label Acply Label	
16 (Ch16 label), OFF 17 (Ch17 label), OFF 19 (Ch10 label), OFF 19 (Ch19 label), OFF	-		

Select "VDC" and "auto" for the range, or the appropriate range for your instrument, and type in the Units in "PSI" or whatever units your transducer will display then click "OK".

Select the Mx+B button. Select "Calculate Mx+B from point values."

Function:	OK
VDC 🔄	Cancel
Anto -	Help
RTD RD	

#### **Current values**

If the value coming in to your data acquisition system is current, you are using a current shunt, and the data acquisition system is reading voltage (4 to 20 mA or other current input). Calculate the minimum and maximum values for the input range using Ohms law ( $E = I \ge R$ ).

**Note:** The voltage value must comply with the limits of the data acquisition system—keep this in mind when choosing a current shunt.

**Example:** I have a 10  $\Omega$  resistor and my transducer is putting 4 to 20 mA into my data acquisition system.

0.004 X 10 = 0.04 V

And

0.020 X 10 = 0.2 V

so now I will enter 0.04 V in the minimum Input Range field and 0.2 V into the maximum input field.

Then I will find the lower and upper display range of the transducer and input that into the Display Range fields.

So if my transducer will measure 0 to 30 psi, I enter 0 in the minimum field and 30 in the maximum field.

Input Range	OK
Minimum 04	Cancel
Masimum: [.2	Help
Display Range	-
Minimum: 0	7
Marine Par	

N. BOLLEY HAVE

olt	age	values
-		

Click OK.

V

If the value coming in to your data acquisition system is voltage (0 to 10 V, 0 to 5 V, etc.), you will configure your software channel like above, except you will simply enter the lower and upper input ranges and the Minimum and Maximum display ranges of the transducer. The data acquisition system is reading straight voltage so you will not have to do any manual calculations.

Click OK. Run a test.

Input Range	OK.
Minimum 0	Cancel
Maximum 20	Help
Display Range	
Hinimum: 0	-
Maximum 30	

Fluke. The Most Trusted Tools in the World.

Fluke Corporation PO Box 9090, Everett, WA 98206 U.S.A.

Fluke Europe B.V. PO Box 1186, 5602 BD Eindhoven, The Netherlands

Modification of this document is not permitted without written permission from Fluke Corporation. For more information call:

In the U.S.A. (800) 443-5853 or Fax (425) 446-5116 In Europe/M-East/Africa +31 (0) 40 2675 200 or Fax +31 (0) 40 2675 222 In Canada (800)-36-FLUKE or Fax (905) 890-6866 From other countries +1 (425) 446-5500 or Fax +1 (425) 446-5116 Web access: http://www.fluke.com

©2012 Fluke Corporation. Specifications subject to change without notice. Printed in U.S.A. 7/2012 4238018A\_EN